



UK DECLARATION OF CONFORMITY

In accordance with UK Government Guidance

Manufacturer

Etac A/S
Parallelvej 3, 8751 Gedved,
Denmark
Phone no. + 45 79685833
www.etac.com

UK Responsible Person

Etac Ltd
Unit 60, Hartlebury Trading Estate,
Hartlebury, Kidderminster, DY10
4JD

Email enquiries.uk@etac.com
Web www.etac.com/uk
Phone +44 (0)121 561 2222

Statement

This declaration of conformity is issued under the sole responsibility of the manufacturer.
The device(s) covered by this declaration of conformity is/are in conformity with UK MDR 2002
(Medical Device Regulations 2002)

Basic UDI-DI

570799541TM

Device description

Ceiling Hoists

GMDN

38129

Intended purpose

The ceiling hoist is an assistive device intended for alleviation of, or compensation for, a functional impairment due to an injury or disability. The device is designed for an individual lacking the ability to stand up and transfer themselves to/from a bed, the floor, a wheelchair, a chair, a toilet or similar due to reduced mobility or physical strength.

Device name(s)

Molift Duo

Risk class of the device

Class I, rule 1

Harmonized/Established Standards

Separate list available upon request

Place


Gedved, Denmark

Date of issue

10. October 2025

Name and function

Sampavi Rajkumar, Quality Specialist



Signature, on behalf of Etac A/S