



# UK DECLARATION OF CONFORMITY

In accordance with UK Government Guidance

<b>Manufacturer</b>	<b>Etac A/S</b> Parallelvej 3 DK-8751 Gedved Denmark  <a href="http://www.etac.com">www.etac.com</a>	<b>UK Responsible Person</b>	<b>Etac Ltd</b> Unit 60, Hartlebury Trading Estate, Hartlebury, Kidderminster, DY10 4JD  Email <a href="mailto:enquiries.uk@etac.com">enquiries.uk@etac.com</a> Web <a href="http://www.etac.com/uk">www.etac.com/uk</a> Phone +44 (0)121 561 2222
<b>Statement</b>	This declaration of conformity is issued under the sole responsibility of the manufacturer. The device(s) covered by this declaration of conformity is/are in conformity with UK MDR 2002 (Medical Device Regulations 2002)		
<b>Basic UDI-DI</b>	57079952EG		
<b>Device description</b>	Sit-To-Stand Aid		
<b>GMDN</b>	37953		
<b>Intended purpose</b>	The sit-to-stand aid is an assistive device intended for alleviation of, or compensation for, a functional impairment due to an injury or disability. The device is designed for an individual lacking the ability to stand up and transfer themselves over shorter distances to another sitting position to/from a bed, a wheelchair, a chair, a toilet or similar due to reduced mobility or physical strength.		
<b>Device name(s)</b>	Molift Raiser Pro		
<b>Risk class of the device</b>	Class I, rule I		
<b>Harmonized/Established Standards</b>	Separate list available upon request		
<b>Place</b>	Gedved, Denmark		
<b>Date of issue</b>	15. January 2024		
<b>Name and function</b>	Michael Bruun, Senior Vice President		

  
\_\_\_\_\_  
Signature, on behalf of Etac A/S